

Soundness Test

BA 215

ASTM C.88 / AASHTO T.104



BA 215.1



BG 021/4



BG 034/2



Soundness Test

Type	Parts Name	Specification	Qty
BA 215.1	Stroke Solution	Sodium sulphate solution (Na ₂ SO ₄)	1 Ltr
BG 021/4	Beaker Glass	1000 ml capacity	1 Pc
BG 034/2	Thermometer	Glass 0 – 100 C	1 Pc

The presence of organic matter and certain chemicals can have a considerable influence on the strength and durability of concrete. The ability of aggregates to resist excessive change in volume (soundness) due to physical changes in the environment is also of importance.

Supplied without :

- BA.204.18 Sieve No. 4
- BA.203.2 Precision Sampler Splitter
- BG.005/1 Triple Beam Balance 2610 gr Cap
- BB.510.9 Water Bath
- BG.016/3 Mixing Bowl
- BG.032/2 Pipete 10 ml
- BG.045/1 Rubber Glove

